

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Issued: 7/10/2015  
Effective: 7/10/2015

Version: DMCIF 2015 R1  
Status: Active

Instrument S/N 100365

Received From Kotzebue PD

Date 9/21/15

Analyst Colleen O'Bryant

1. Reason for instrument return:

Software issues - unable to load ticket db,  
followed by a loop of pop ups and restarts.

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Upload  unable

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs  N/A

From NPAS

7. NPAS Repair Paperwork (attached)