

# Alaska Scientific Crime Detection Laboratory

## DataMaster Calibration Form

Issued: 6/11/2014  
Effective: 6/11/2014

Version: DMCAF 2014 R0  
Status: Active

Instrument # 100687

Date: 8/12/15

### Reason for Calibration

SI board was replaced, calibration lost.

#### 1. Wet Bath Simulator Solution

- a. Level 0.100
- b. Manufacturer Guth
- c. Lot 15050
- d. Expiration Date 3/9/17

#### 2. Wet Bath Simulator

- a. Serial Number 501200
- b. Date Calibrated 1/20/15

#### 3. Attach Calibration Factors

CALIBRATION FACTORS

-----  
Alaska Department of Public Safety  
DATAMASTER dmt: 100687  
-----

Date: 08/12/2015  
Time: 13:14:02

OPERATOR NAME:  
COLLEEN S OBRYANT  
CERTIFICATE #: 6237

Ca = 0.1000  
CAL = 0.976974    0.800 <= CAL < 1.200  
b1 = 0.0007    0.0000 <= b1 < 0.0040  
b2 = 0.0055    0.0010 <= b2 < 0.0100  
b3 = 0.0003    0.0000 <= b3 < 0.0040  
Xq = 0.1360    0.0500 <= Xq < 0.2000  
a21 = 1.147831    1.050 <= a21 < 1.300  
a31 = 0.531572    0.300 <= a31 < 0.800