

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 2/24/2026

Instrument #: 100366

Instrument Location: Kodiak Jail Datamaster Room

Supervisor Name and Agency: Phillip Christman with the Kodiak Police Department

Describe the Status Change or Issue with the Instrument:

The instrument was unable to send a VOC. When trying to pull the files from the DMT, was unable to procure it with the flashdrive. Received new instrument on 02/24/2026 at approximately 1330 hours.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: [Click or tap here to enter text.](#)

Out of Service Date/Time: 02/24/2026 at approximately 1400 hours.

-----  
**For Use by SCDL**

Additional Notes

Data file reconciliation is needed upon return to the laboratory. DJW 02/25/2026

**Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov).**

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor