

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 12/21/2022

Instrument #: 100668

Instrument Location: Anchorage Jail – South processing room

Supervisor Name and Agency: Jason Cusack, Anchorage Police

Describe the Status Change or Issue with the Instrument:

On 12/16/2022, I received an email from Carlie with the crime lab requesting I perform a tank change on this instrument as it did not pass a VOC and it seemed like there was some sort of interference with the dry gas tank. I was on leave and saw the email on 12/20/2022 and notified her that I would perform the tank change on 12/21/2022 when I worked next. I performed the tank change at about 2104 hours and the instrument passed VOC immediately after. I also performed a non-drinking subject sample and it worked normally.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: [Click or tap here to enter text.](#)

Out of Service Date/Time: [Click or tap here to enter text.](#)

For Use by SCDL

Additional Notes

[Click or tap here to enter text.](#)

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

Glaister, Carlie K (DPS)

From: Cusack, Jason <jason.cusack@anchorageak.gov>
Sent: Wednesday, December 21, 2022 9:21 PM
To: Glaister, Carlie K (DPS)
Subject: RE: 100668
Attachments: Change-in-Instrument-Status-Form 100668 12-21-2022.docx

I also discovered we only have one spare alco bottle, going into the holidays is there a chance we can get another two or three bottles either on Thursday, Friday, or sometime next week?

Jason

From: Glaister, Carlie K (DPS) <carlie.glaister@alaska.gov>
Sent: Tuesday, December 20, 2022 10:49 AM
To: Cusack, Jason <jason.cusack@anchorageak.gov>
Subject: RE: 100668

[EXTERNAL EMAIL]

Sounds great, thank you.

Carlie Glaister (Bailey)

Forensic Scientist III – Chemistry Section
State of Alaska Scientific Crime Detection Laboratory
907-269-5689

From: Cusack, Jason <jason.cusack@anchorageak.gov>
Sent: Tuesday, December 20, 2022 10:36 AM
To: Glaister, Carlie K (DPS) <carlie.glaister@alaska.gov>
Subject: RE: 100668

CAUTION: This email originated from outside the State of Alaska mail system. Do not click links or open attachments unless you recognize the sender and know the content is safe.

I have been out of town but will check this out tomorrow night when I go back to work next.

Jason

From: Glaister, Carlie K (DPS) <carlie.glaister@alaska.gov>
Sent: Friday, December 16, 2022 10:30 AM
To: Cusack, Jason <jason.cusack@anchorageak.gov>
Subject: 100668

[EXTERNAL EMAIL]

Hello. I am working on VOC day and APD's instrument 100668 has not managed to get a passing VOC. It seems to be having some issues with tank interference. Can you please try installing a new tank to see if this helps? Please send me a change in instrument status form to document this.

Thank you for your help,
Carlie Glaister