

In Field Instrument Review for 4/12/2017 to 5/30/2017

100255(Clear AFB): RFI 2/8- No action taken

100345(AST/Soldotna): RFI 2/24- No action taken

100356(Skagway PD): Pump error 1/9, Blank error 1/9- No action taken

100361(Unalaska PD): RFI 2/28, Pump error 1/28, Interference 1/28, Tank Interference 2/28, Blank Error 3/28- Also had a blank error on the 5/30/2017 VOC- Contact Unalaska about instrument location and ventilation

100363(Nome PD): Interference 1/8- No action taken

100364(APD): Blank Error 1/104, Gas Flow Error 2/104, Filter Wheel Error 1/104- No action taken

100365(North Pole PD): Pump Error 1/20, Interference 2/20, Tank Interference 1/20- No action taken

100370(Homer PD): RFI 2/16, Interference 2/16, Tank Interference 1/16, Internal Standard Error 1/16- Watch instrument, Instrument will be replaced prior to PM due in October unless further problems observed.

100374(Hoonah PD): Removed from Service- Touchscreen not responding

100378(AST/Fairbanks): RFI 1/37, Gas Flow Error 1/37- No action taken

100387(Fairbanks PD): RFI 2/44, Blank Error 1/44- No action taken

100396(APD): RFI 2/125, Interference 1/125, Tank Interference 2/125, Internal Standard Error 1/125- No action taken

100397(AST/Fairbanks) : RFI 1/23- Removed from Service to Troubleshoot modem line issue

100398(Kotzebue PD): Interference 2/12, Blank Error 3/12, Gas Flow Error 1/12, Ambient Fail 1/12- Replace instrument

100400(NSBPD): Blank Error 1/9, Standard Out of Range 1/9- Replacement external standard sent

100402(Anchorage Airport): Blank Error 1/7- Removed from Service 5/19

100413(AST/Cooper Landing): Blank Error 1/3- No action taken

100414(Eielson AFB): Blank Error 1/7- No action taken

100421(JBER): RFI 5/61, Pump Error 3/61, Interference 7/61, Tank Interference 1/61, Blank Error 8/61- troubleshooting for environmental issues.

100666(Bethel PD): RFI 1/13- No action taken

100669(APD): RFI 1/24, Pump error 1/24, Interference 3/24, Blank Error 1/24: No action taken

100682(Hooper Bay PD): Pump Error 1/3- No action taken

100690(Sitka PD): Interference 3/33, Tank Interference 3/33, Blank Error 8/33- removed from service
6/1/2017

100691(Valdez PD): RFI 3/10- No action taken

100695(Wasilla PD): RFI 2/37- No action taken

100698(Kenai PD): Tank Interference 1/23, Blank Error 1/23- No action taken

100700(Soldotna PD): RFI 1/8- No action taken

